



# Certificate of Compliance

**Certificate:** 70202313

**Master Contract:** 205078

**Project:** 80161739

**Date Issued:** 2023-03-29

**Issued To:** Analog Devices Inc.  
804 Woburn St  
Wilmington, Massachusetts, 01887-3462  
United States

**Attention:** Jason Naso

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicator ▲*



**Issued by:** *Martin Buchanan*  
Martin Buchanan, P. Eng.

**PRODUCTS**

CLASS - C907330 - ELECTRONIC COMPONENTS Optoisolators and non-optical isolating devices

Component Acceptance of Optoisolator Like Magnetic Coupling Devices:

Device	Rating		Standard/Notice and Clauses	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
(SOIC 16W RW package) ADuM5020-5BRWZ ADuM5020-3BRWZ ADuM5020-5WBRWZ ADuM5020-3WBRWZ	3.0	125	CSA 14-18 tb35, 6.21.4.1, 6.2.1/6.2.12, 6.8.1 62368-1:19, UPD1:2021 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.8, 5.4.1.5.3, 5.4.9.1, 5.4.1.4 61010-1-12+A1 6.7.1.3, 6.7.2.1 or tbK.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10	-	-	7.8



**Certificate:** 70202313  
**Project:** 80161739

**Master Contract:** 205078  
**Date Issued:** 2023-03-29

Device	Rating		Standard/Notice and Clauses	Internal		External
	kV	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
(SOIC 8W RI-8-1 package) ADuM5028-5BRIZ ADuM5028-3BRIZ ADuM5028-5WBRIZ ADuM5028-3WBRIZ	3.0		60601-1:14 (R2018) 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7  IEC 62368-1:2018 Ed. 3 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.8, 5.4.1.5.3, 5.4.9.1, 5.4.1.4 61010-1 3 <sup>rd</sup> Ed+A1 6.7.1.3, 6.7.2.1 or tbK.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10 60601-1:2005 Ed 3.0+A1+A2 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7  EN 62368-1:2020+A11:2020 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.8, 5.4.1.5.3, 5.4.9.1, 5.4.1.4			8.3
(SOIC 16W RI-16-2 package) ADuM6020-5BRIZ ADuM6020-3BRIZ ADuM6020-5WBRIZ ADuM6020-3WBRIZ  (SOIC 8W RI-8-1 package) ADuM6028-5BRIZ ADuM6028-3BRIZ ADuM6028-5WBRIZ ADuM6028-3WBRIZ	5.0					

where Z if used indicates the lead-free version. Suffix EP at the end of the base part number indicates lower extended temperature capability to -55C with a NiPdAu lead finish. Further suffix letters or digits differentiate shipping package formats.

Notes:

1. The RW devices meet basic insulation requirements for 780Vrms and the RI devices meet basic insulation requirements for 830Vrms for CSA 62368-1:19, UPD1:2021, IEC 62368-1:2018 Ed. 3 and EN 62368-1:2020+A11:2020. (pollution degree 2, material group III)
2. The 5kV RI devices meet basic insulation requirements for 830Vrms and the 3kV RI and the RW devices meet basic insulation requirements for 600Vrms including 400/690Vrms for CSA 62368-1:19, UPD1:2021, IEC 62368-1:2018 Ed. 3 and EN 62368-1:2020+A11:2020. (pollution degree 2, material group III)
3. The RW devices meet reinforced insulation requirements for 390Vrms and the RI devices meet reinforced insulation requirements for 415Vrms for CSA 62368-1:19, UPD1:2021, IEC 62368-1:2018 Ed. 3 and EN 62368-1:2020+A11:2020. (pollution degree 2, material group III)
4. The 5kV RI devices meet reinforced insulation requirements for 415Vrms and the 3kV RI devices meet 300Vrms including 230/400Vrms and 277/480Vrms (limited to 415Vrms) and the RW devices meet reinforced insulation requirements for 300Vrms including 230/400Vrms and 277/480Vrms (limited to 390Vrms) for CSA 62368-1:19, UPD1:2021, IEC 62368-1:2018 Ed. 3 and EN 62368-1:2020+A11:2020. (pollution degree 2, material group III)



**Certificate:** 70202313  
**Project:** 80161739

**Master Contract:** 205078  
**Date Issued:** 2023-03-29

5. For CSA 61010-1-12+A1 and IEC 61010-1 3<sup>rd</sup> Ed.+A1 the devices meet 600Vrms for basic insulation and 300V for reinforced insulation based on 61010-1 Cl 14.1 a) for use in 61010-1 end products because they meet the requirements of the 62368-1 evaluation. (pollution degree 2, material group III)
6. For CSA 60601-1:14 (R2018) and IEC60601-1 Ed. 3+A1+A2 for 1 MOPP for 419Vrms, the RW and 3kV RI devices meet clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.
7. For CSA 60601-1:14 (R2018) and IEC60601-1 Ed. 3+A1+A2 for 1 MOPP for 590Vrms, the 5kV RI devices meet clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.
9. Case material: material group I.
10. Evaluated by thermal cycling and other tests for a temperature rating of 125C.
11. The creepage and clearance has been evaluated for altitudes  $\leq 2000\text{m}$ , in pollution degree 2 and overvoltage category II except where specified above. (pollution degree 2, material group III).

These devices are Component Accepted as components for use in other Certified equipment where the suitability of the combination shall be determined by investigation in the final application.

### **APPLICABLE REQUIREMENTS**

- CAN/CSA-C22.2 No 14-18 - Industrial Control Equipment
- CAN/CSA-C22.2 No. 62368-1:19, UPD1:2021 - Audio/video, information and communication technology equipment - Part 1: Safety requirements (IEC 62368-1:2018, MOD)
- IEC 62368-1:2018 Ed. 3 - Audio/video, information and communication technology equipment - Part 1: Safety requirements
- EN 62368-1:2020+A11:2020 - Audio/video, information and communication technology equipment - Part 1: Safety requirements (IEC 62368-1:2018)

Clauses 6.7.1.3, 6.7.2.1 or K.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10 of

CAN/CSA C22.2 No. 61010-1-12, UPD1: 2015, UPD2: 2016, AMD1: 2018 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Tri-national standard, with UL 61010-1 Ed. 3 (2012), AMD1: 2018 and ANSI/ISA-61010-1 (82.02.01))

IEC 61010-1:2010 Ed. 3.1:2017 01 - Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 1: General Requirements

Clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 of

CAN/CSA C22.2 60601-1:14 (R2018) - Medical Electrical Equipment Part 1: General requirements for basic safety and general performance (Adopted IEC60601-1:2005 Edition 3.0 +Amendment 1, 2012-07 MOD +Amendment 2, 2020-08 MOD)

and



**Certificate:** 70202313  
**Project:** 80161739

**Master Contract:** 205078  
**Date Issued:** 2023-03-29


---

IEC60601-1:2005 Ed 3.0+A1+A2 - Medical Electrical Equipment Part 1: General requirements for basic safety and general performance

Notes:

---

Products certified under Class C907330 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). [www.scc.ca](http://www.scc.ca)





## *Supplement to Certificate of Compliance*

**Certificate:** 70202313

**Master Contract:** 205078

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80161739	2023-03-29	Update 70202313 to 62368-1 Ed 3, 60601-1 A2 and correct moulding compound to EME-G770H Type CD.
70202313	2019-01-29	Original Component Acceptance