



# Certificate of Compliance

**Certificate:** 80020409

**Master Contract:** 205078

**Project:** 80020409

**Date Issued:** 2020-04-30

**Issued To:** Analog Devices Inc.  
804 Woburn St  
Wilmington, Massachusetts, 01887-3462  
United States

**Attention:** Mark Cantrell



**Issued by:** *Martin Buchanan*  
Martin Buchanan, P. Eng.

## PRODUCTS

CLASS - C907330 - ELECTRONIC COMPONENTS Optoisolators and non-optical isolating devices

Component Acceptance of Optoisolator Like Devices:

Device	Ratings		Standard/Notice and Clauses	Internal		External
	(kV)	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
LGA 20 ADE1201ACCZ ADE1202ACCZ	3.75	125	CSA 14-18 tb35, 6.2.1.1, 6.2.1.1/6.2.10, 6.8.1.1 60950-1-07+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1-14 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 5.4.9.1, 5.4.8, 5.4.1.4, 5.4.3.3 61010-1-12+A1 6.7.1.3, 6.7.2.1 or tbK.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10 60601-1:14 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7  IEC	-	-	6.8



**Certificate:** 80020409  
**Project:** 80020409

**Master Contract:** 205078  
**Date Issued:** 2020-04-30

Device LGA 20	Ratings		Standard/Notice and Clauses	Internal		External
	(kV)	°C		Creepage (mm)	Dist Thru (mm)	Creep/Clear (mm)
			60950-1 2nd Ed.,+A1+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1:2014 Ed. 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 5.4.9.1, 5.4.8, 5.4.1.4, 5.4.3.3 61010-1 3 <sup>rd</sup> Ed+A1 6.7.1.3, 6.7.2.1 or tbK.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10 60601-1 Ed.3+A1 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7  EN 60950-1:2006+A1+A12+A2 2.10.3.3, 2.10.4.2, 2.10.4.3, 2.10.5.4a, 2.10.11, 4.5.2, 5.2 62368-1:2014/A11:2017 5.4.3, 5.4.2, 5.4.4.4, 5.4.7, 5.4.1.5.3, 5.4.9.1, 5.4.8, 5.4.1.4, 5.4.3.3			

where Z if used indicates the lead-free version. Suffix EP at the end of the base part number indicates lower extended temperature capability to -55C with a NiPdAu lead finish. Further suffix letters or digits differentiate shipping package formats.

**Notes:**

1. The devices meet basic insulation requirements for 680Vrms and reinforced insulation requirements for 340Vrms for CSA 60950-1-07+A1+A2, CSA 62368-1-14, IEC 60950-1 2<sup>nd</sup> Ed.+A1+A2, IEC 62368-1:2014 and EN60950-1:2006+A1+A12+A2 and EN62368-1:2014/A11:2017. (pollution degree 2, material group III)
2. For CSA 61010-1-12+A1 and IEC 61010-1 3<sup>rd</sup> Ed. the devices meet 600Vrms for basic insulation for overvoltage category III.
3. For CSA 61010-1-12+A1 and IEC 61010-1 3<sup>rd</sup> Ed. the devices meet 300V for reinforced insulation based on 61010-1 Cl 14.1 a) for use in 61010-1 end products because they meet the requirements of the 62368-1 evaluation. The risk management process is not applicable to these clauses. (pollution degree 2, material group III)
4. For CSA 60601-1:14 and IEC60601-1 Ed.3+A1 for 2 MOPP for 175Vrms, the devices meet clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7. The risk management process is not applicable to these clauses.
5. Case material: material group I.
6. Evaluated by thermal cycling and other tests for a temperature rating of 125C.
7. The creepage and clearance has been evaluated for altitudes ≤ 2000m, in pollution degree 2 and overvoltage category II except where specified above. (pollution degree 2, material group III).

These devices are Component Accepted as components for use in other Certified equipment where the suitability of the combination shall be determined by investigation in the final application.



**Certificate:** 80020409  
**Project:** 80020409

**Master Contract:** 205078  
**Date Issued:** 2020-04-30

### APPLICABLE REQUIREMENTS

- CAN/CSA-C22.2 No 14-18 - Industrial Control Equipment  
CAN/CSA-C22.2 No 60950-1-07+A1+A2 - Information Technology Equipment - Safety - Part 1: General Requirements (Bi-national Standard, with UL 60950-1)  
CAN/CSA-C22.2 No. 62368-1-14 - Audio/video, information and communication technology equipment– Part 1: Safety requirements  
IEC 60950-1 2<sup>nd</sup> Ed.+A1+A2 - Information Technology Equipment - Safety - Part 1: General Requirements  
IEC 62368-1:2014 Ed. 2 - Audio/video, information and communication technology equipment– Part 1: Safety requirements  
EN 60950-1:2006+A1+A12+A2 - Information Technology Equipment - Safety - Part 1: General Requirements  
EN 62368-1:2014/A11:2017 - Audio/video, information and communication technology equipment - Part 1: Safety requirements (IEC 62368-1:2014, modified)

Clauses 6.7.1.3, 6.7.2.1 or K.1 to K.4, 6.7.2.2.1, 6.7.2.2.2 or tbK.9, A.17, tbK.5x1.6, K.6x1.6, K.7x1.6, 10 of

CAN/CSA C22.2 No. 61010-1-12, UPD1: 2015, UPD2: 2016, AMD1: 2018 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Tri-national standard, with UL 61010-1 Ed. 3 (2012), AMD1: 2018 and ANSI/ISA-61010-1 (82.02.01))

IEC 61010-1:2010 Ed. 3.1:2017 01 - Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 1: General Requirements

Clauses 8.5.5.1, 8.8.2, 8.8.3x1.6, 8.9.3.2, 8.9.3.4, 8.9.1.15, 8.9.1.7 of

CAN/CSA C22.2 60601-1:14 - Medical Electrical Equipment Part 1: General requirements for basic safety and general performance (Adopted IEC60601-1:2005 Edition 3.0 +Amendment 1, 2012-07, MOD)

and

IEC60601-1:2005 Ed 3.0+A1 - Medical Electrical Equipment Part 1: General requirements for basic safety and general performance



## *Supplement to Certificate of Compliance*

**Certificate:** 80020409

**Master Contract:** 205078

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80020409	2020-04-30	Original Component Acceptance